

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031320\
 Data File : FF006550.D
 Signal(s) : FID2B.ch
 Acq On : 13 Mar 2020 10:59
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 13 23:04:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031020.M
 Quant Title :
 QLast Update : Tue Mar 10 14:59:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.586	6778091	52.665 ug/ml
Target Compounds			
2) N-DECANE	4.358	7786991	56.572 ug/ml
3) N-DODECANE	6.372	7954895	56.401 ug/ml
4) N-TETRADECANE	8.162	7951705	55.600 ug/ml
5) N-HEXADECANE	9.757	8049665	54.914 ug/ml
6) N-OCTADECANE	11.193	8037837	54.396 ug/ml
7) N-EICOSANE	12.497	7959670	53.555 ug/ml
8) N-DOCOSANE	13.690	7859222	52.917 ug/ml
10) N-TETRACOSANE	14.789	7731051	52.309 ug/ml
11) N-HEXACOSANE	15.805	7521333	51.901 ug/ml
12) N-OCTACOSANE	16.750	7414947	51.771 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031320\
Data File : FF006550.D
Signal(s) : FID2B.ch
Acq On : 13 Mar 2020 10:59
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 13 23:04:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031020.M
Quant Title :
QLast Update : Tue Mar 10 14:59:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

