

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031420\
 Data File : FF006566.D
 Signal(s) : FID2B.ch
 Acq On : 14 Mar 2020 10:14
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 16 00:51:04 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031020.M
 Quant Title :
 QLast Update : Tue Mar 10 14:59:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.588	6146825	47.760 ug/ml
Target Compounds				
2)	N-DECANE	4.359	7230189	52.527 ug/ml
3)	N-DODECANE	6.374	7340684	52.046 ug/ml
4)	N-TETRADECANE	8.163	7307266	51.094 ug/ml
5)	N-HEXADECANE	9.758	7380102	50.346 ug/ml
6)	N-OCTADECANE	11.195	7328943	49.598 ug/ml
7)	N-EICOSANE	12.499	7259871	48.846 ug/ml
8)	N-DOCOSANE	13.693	7139191	48.069 ug/ml
10)	N-TETRACOSANE	14.790	6999178	47.357 ug/ml
11)	N-HEXACOSANE	15.806	6791668	46.866 ug/ml
12)	N-OCTACOSANE	16.752	6681647	46.651 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031420\
Data File : FF006566.D
Signal(s) : FID2B.ch
Acq On : 14 Mar 2020 10:14
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 16 00:51:04 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031020.M
Quant Title :
QLast Update : Tue Mar 10 14:59:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

