

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031420\
 Data File : FF006575.D
 Signal(s) : FID2B.ch
 Acq On : 14 Mar 2020 15:17
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 16 00:52:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031020.M
 Quant Title :
 QLast Update : Tue Mar 10 14:59:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.590	6049134	47.001 ug/ml
Target Compounds				
2)	N-DECANE	4.360	7341871	53.338 ug/ml
3)	N-DODECANE	6.375	7456437	52.867 ug/ml
4)	N-TETRADECANE	8.165	7398076	51.729 ug/ml
5)	N-HEXADECANE	9.760	7460629	50.896 ug/ml
6)	N-OCTADECANE	11.196	7390558	50.015 ug/ml
7)	N-EICOSANE	12.501	7293318	49.071 ug/ml
8)	N-DOCOSANE	13.694	7153717	48.167 ug/ml
10)	N-TETRACOSANE	14.793	6999220	47.358 ug/ml
11)	N-HEXACOSANE	15.807	6770576	46.720 ug/ml
12)	N-OCTACOSANE	16.754	6659918	46.499 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031420\
Data File : FF006575.D
Signal(s) : FID2B.ch
Acq On : 14 Mar 2020 15:17
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 16 00:52:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031020.M
Quant Title :
QLast Update : Tue Mar 10 14:59:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

