

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031423\
 Data File : FF012266.D
 Signal(s) : FID2B.ch
 Acq On : 14 Mar 2023 15:47
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/15/2023
 Supervised By :Ankita Jodhani 03/15/2023

Integration File: autoint1.e
 Quant Time: Mar 15 04:30:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF022123.M
 Quant Title :
 QLast Update : Tue Feb 21 16:12:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.028	5216942	43.972 ug/mlm
Target Compounds			
2) N-DECANE	4.591	5904679	44.035 ug/ml
3) N-DODECANE	6.753	6052197	44.547 ug/ml
4) N-TETRADECANE	8.579	6154178	44.207 ug/ml
5) N-HEXADECANE	10.184	6288141	44.174 ug/ml
6) N-OCTADECANE	11.625	6298579	44.307 ug/ml
7) N-EICOSANE	12.933	6214193	44.046 ug/ml
8) N-DOCOSANE	14.130	5930201	43.962 ug/ml
10) N-TETRACOSANE	15.232	5807588	43.836 ug/ml
11) N-HEXACOSANE	16.252	5588535	43.810 ug/ml
12) N-OCTACOSANE	17.201	5397542	43.976 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031423\
Data File : FF012266.D
Signal(s) : FID2B.ch
Acq On : 14 Mar 2023 15:47
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/15/2023
Supervised By :Ankita Jodhani 03/15/2023

Integration File: autoint1.e
Quant Time: Mar 15 04:30:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF022123.M
Quant Title :
QLast Update : Tue Feb 21 16:12:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

