

rteres

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF03142
Data File : FF012266.D
Signal(s) : FID2B.ch
Acq On : 14 Mar 2023 15:47
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/15/2023
Supervised By :Ankita Jodhani 03/15/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF022123.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.591	4.558	4.678	BB	580670	5904679	93.75%	9.107%
2	6.753	6.717	6.874	BB	574568	6052197	96.09%	9.334%
3	8.579	8.533	8.688	BB	568533	6154178	97.71%	9.492%
4	10.184	10.144	10.285	BB	544152	6288141	99.83%	9.698%
5	11.625	11.583	11.719	BB	525625	6298579	100.00%	9.714%
6	12.933	12.893	13.011	BB	512084	6214193	98.66%	9.584%
7	14.130	14.094	14.197	VB	491848	5930201	94.15%	9.146%
8	15.028	14.976	15.058	BV	384886	5201380	82.58%	8.022%
9	15.232	15.180	15.293	BB	452918	5807588	92.20%	8.957%
10	16.252	16.197	16.312	BB	418809	5588535	88.73%	8.619%
11	17.201	17.132	17.258	BB	392969	5397542	85.69%	8.325%
Sum of corrected areas:						64837213		

FF022123.M Wed Mar 15 04:56:38 2023