

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031622\
 Data File : FF010029.D
 Signal(s) : FID2B.ch
 Acq On : 16 Mar 2022 16:07
 Operator : AJ\MA
 Sample : PB143390BS
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 PB143390BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/17/2022
 Supervised By :Ankita Jodhani 03/17/2022

Integration File: autoint1.e
 Quant Time: Mar 17 00:34:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031022.M
 Quant Title :
 QLast Update : Thu Mar 10 12:46:01 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.694	3146263	20.979 ug/mlm
Target Compounds			
1) N-OCTANE	2.337	3106606	18.744 ug/ml
2) N-DECANE	4.436	3727535	21.856 ug/ml
3) N-DODECANE	6.454	3867205	22.490 ug/ml
4) N-TETRADECANE	8.247	3994782	23.008 ug/ml
5) N-HEXADECANE	9.847	4027264	23.121 ug/ml
6) N-OCTADECANE	11.288	4128306	23.089 ug/ml
7) N-EICOSANE	12.597	4119352	23.723 ug/ml
8) N-DOCOSANE	13.796	4033753	23.593 ug/ml
10) N-TETRACOSANE	14.899	4017145	23.785 ug/ml
11) N-HEXACOSANE	15.920	3890706	23.562 ug/ml
12) N-OCTACOSANE	16.870	3766323	23.532 ug/ml
13) N-TRIACONTANE	17.758	3571936	23.255 ug/ml
14) N-DOTRIACONTANE	18.591	3184375	21.805 ug/ml
15) N-TETRATRIACONTANE	19.373	2629016	19.195 ug/mlm
16) N-HEXATRIACONTANE	20.114	2087950	16.792 ug/ml
17) N-OCTATRIACONTANE	20.852	1754314	15.754 ug/mlm
18) N-TETRACONTANE	21.763	1449674	14.127 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031622\
Data File : FF010029.D
Signal(s) : FID2B.ch
Acq On : 16 Mar 2022 16:07
Operator : AJ\MA
Sample : PB143390BS
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
PB143390BS

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 03/17/2022
Supervised By : Ankita Jodhani 03/17/2022

Integration File: autoint1.e
Quant Time: Mar 17 00:34:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031022.M
Quant Title :
QLast Update : Thu Mar 10 12:46:01 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

