

rteres

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF03172
Data File : FF010054.D
Signal(s) : FID2B.ch
Acq On : 17 Mar 2022 13:52
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Reviewed By :Yogesh Patel 03/18/2022
Supervised By :Ankita Jodhani 03/18/2022

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031022.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.334	2.301	2.381	BB	1099768	7878552	92.98%	5.802%
2	4.435	4.401	4.495	BB	890457	8070680	95.24%	5.944%
3	6.454	6.415	6.505	BB	829117	8131481	95.96%	5.988%
4	8.248	8.211	8.311	BB	793911	8224679	97.06%	6.057%
5	9.848	9.798	9.888	BB	766394	8260371	97.48%	6.083%
6	11.290	11.228	11.325	BV	736732	8473816	100.00%	6.240%
7	12.599	12.558	12.658	BB	685447	8202267	96.80%	6.040%
8	13.797	13.759	13.838	VB	665396	8079368	95.35%	5.950%
9	14.696	14.655	14.808	BB	520437	7092326	83.70%	5.223%
10	14.901	14.831	14.968	BB	621868	7978307	94.15%	5.876%
11	15.922	15.825	15.975	BB	587842	7774651	91.75%	5.726%
12	16.872	16.688	16.908	BB	560600	7572582	89.36%	5.577%
13	17.761	17.691	17.801	BB	537043	7352083	86.76%	5.414%
14	18.594	18.521	18.638	BB	528913	7147015	84.34%	5.263%
15	19.377	19.068	19.415	BV	486214	6974472	82.31%	5.136%
16	20.116	20.061	20.155	BV	443692	6591031	77.78%	4.854%
17	20.854	20.797	20.900	M	339412	6024937	71.10%	4.437%
18	21.766	21.645	21.865	BB	249741	5960515	70.34%	4.390%
Sum of corrected areas:						135789134		

FF031022.M Fri Mar 18 11:27:31 2022