

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031822\
 Data File : FF010089.D
 Signal(s) : FID2B.ch
 Acq On : 18 Mar 2022 23:07
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: Sample.e
 Quant Time: Mar 19 01:21:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031022.M
 Quant Title :
 QLast Update : Thu Mar 10 12:46:01 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.695	6157754	41.060 ug/ml
Target Compounds			
2) N-DECANE	4.434	7759621	45.497 ug/ml
3) N-DODECANE	6.452	7757005	45.111 ug/ml
4) N-TETRADECANE	8.245	7733172	44.539 ug/ml
5) N-HEXADECANE	9.845	7623269	43.766 ug/ml
6) N-OCTADECANE	11.287	7703788	43.085 ug/ml
7) N-EICOSANE	12.597	7351965	42.340 ug/ml
8) N-DOCOSANE	13.795	7120805	41.648 ug/ml
10) N-TETRACOSANE	14.899	6929679	41.029 ug/ml
11) N-HEXACOSANE	15.919	6683749	40.477 ug/ml
12) N-OCTACOSANE	16.869	6491166	40.557 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031822\
Data File : FF010089.D
Signal(s) : FID2B.ch
Acq On : 18 Mar 2022 23:07
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: Sample.e
Quant Time: Mar 19 01:21:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031022.M
Quant Title :
QLast Update : Thu Mar 10 12:46:01 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

