

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032023\
 Data File : FF012304.D
 Signal(s) : FID2B.ch
 Acq On : 20 Mar 2023 09:29
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/21/2023
 Supervised By :Ankita Jodhani 03/21/2023

Integration File: autoint1.e
 Quant Time: Mar 21 03:31:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031623.M
 Quant Title :
 QLast Update : Wed Mar 15 18:07:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.026	5416662	48.401 ug/ml
Target Compounds			
1) N-OCTANE	2.076	5954869	46.958 ug/ml
2) N-DECANE	4.593	6316920	50.454 ug/ml
3) N-DODECANE	6.755	6412612	51.728 ug/ml
4) N-TETRADECANE	8.580	6476239	49.700 ug/ml
5) N-HEXADECANE	10.184	6536120	48.981 ug/ml
6) N-OCTADECANE	11.625	6554704	48.797 ug/ml
7) N-EICOSANE	12.933	6449406	48.482 ug/ml
8) N-DOCOSANE	14.127	6142573	48.339 ug/ml
10) N-TETRACOSANE	15.230	6017835	48.255 ug/ml
11) N-HEXACOSANE	16.248	5787267	48.332 ug/ml
12) N-OCTACOSANE	17.197	5580101	48.510 ug/ml
13) N-TRIACONTANE	18.081	5342459	48.553 ug/ml
14) N-DOTRIACONTANE	18.913	4979663	47.981 ug/ml
15) N-TETRATRIACONTANE	19.694	4573556	48.343 ug/ml
16) N-HEXATRIACONTANE	20.432	4168148	47.576 ug/ml
17) N-OCTATRIACONTANE	21.165	4045094	48.855 ug/ml
18) N-TETRACONTANE	22.074	3962176	48.504 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032023\
Data File : FF012304.D
Signal(s) : FID2B.ch
Acq On : 20 Mar 2023 09:29
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/21/2023
Supervised By :Ankita Jodhani 03/21/2023

Integration File: autoint1.e
Quant Time: Mar 21 03:31:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031623.M
Quant Title :
QLast Update : Wed Mar 15 18:07:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

