

rteres

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF03202
Data File : FF012313.D
Signal(s) : FID2B.ch
Acq On : 20 Mar 2023 14:37
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/21/2023
Supervised By :Ankita Jodhani 03/21/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031623.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.074	2.035	2.150	BB	515828	5507258	89.01%	5.677%
2	4.593	4.561	4.660	BB	588399	5924774	95.76%	6.107%
3	6.756	6.721	6.836	BB	588200	6001662	97.01%	6.186%
4	8.582	8.544	8.665	BB	570772	6067413	98.07%	6.254%
5	10.187	10.150	10.260	BB	563353	6186897	100.00%	6.377%
6	11.628	11.588	11.698	BB	531929	6170357	99.73%	6.360%
7	12.937	12.887	13.012	BB	494966	6100359	98.60%	6.288%
8	14.133	14.083	14.204	BB	467372	5830820	94.24%	6.010%
9	15.030	14.978	15.103	BB	363076	5144401	83.15%	5.303%
10	15.234	15.182	15.303	BB	453416	5722725	92.50%	5.899%
11	16.253	16.199	16.326	BB	425670	5508961	89.04%	5.678%
12	17.202	17.135	17.269	BB	380784	5345844	86.41%	5.510%
13	18.087	18.026	18.145	BB	361109	5148023	83.21%	5.306%
14	18.918	18.860	18.970	BB	335228	4868732	78.69%	5.018%
15	19.700	19.550	19.745	BB	313653	4537238	73.34%	4.677%
16	20.440	20.100	20.488	PV	287783	4178254	67.53%	4.307%
17	21.174	21.042	21.251	BB	238699	4371351	70.65%	4.506%
18	22.086	22.000	22.158	BB	182817	4402475	71.16%	4.538%
Sum of corrected areas:						97017543		

FF031623.M Tue Mar 21 03:55:56 2023