

rters

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/28/2022
Supervised By :mohammad ahmed 03/30/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032622
Data File : FF010173.D
Signal(s) : FID2B.ch
Acq On : 27 Mar 2022 01:09
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF032622.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.280	2.245	2.338	BB	929526	6907765	98.07%	6.622%
2	4.371	4.328	4.428	BB	777902	7043716	100.00%	6.753%
3	6.390	6.352	6.448	BB	718790	7012453	99.56%	6.723%
4	8.185	8.148	8.248	BB	671141	6990950	99.25%	6.702%
5	9.787	9.745	9.855	BB	626597	6885446	97.75%	6.601%
6	11.229	11.162	11.262	BB	588478	6879261	97.67%	6.595%
7	12.540	12.498	12.602	BB	549284	6468206	91.83%	6.201%
8	13.740	13.662	13.782	BB	487785	6165528	87.53%	5.911%
9	14.640	14.595	14.745	BB	385563	5311927	75.41%	5.092%
10	14.844	14.765	14.908	BB	469983	5902603	83.80%	5.659%
11	15.866	15.802	15.918	BB	421383	5638749	80.05%	5.406%
12	16.817	16.732	16.882	BB	397777	5469248	77.65%	5.243%
13	17.705	17.632	17.748	BB	373946	5344449	75.88%	5.123%
14	18.539	18.345	18.583	BV	372525	5214371	74.03%	4.999%
15	19.322	19.138	19.362	PV	328164	4860174	69.00%	4.659%
16	20.062	20.012	20.090	BBA	292867	4192525	59.52%	4.019%
17	20.791	20.712	20.833	BV	222525	3849187	54.65%	3.690%
18	21.681	21.572	21.845	BV	173799	4175964	59.29%	4.003%
Sum of corrected areas:						104312520		

FF032622.M Tue Mar 29 12:59:25 2022