

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032719\
 Data File : FF004000.D
 Signal(s) : FID2B.ch
 Acq On : 27 Mar 2019 13:06
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Mar 28 00:40:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030119.M
 Quant Title :
 QLast Update : Mon Mar 04 08:35:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.573	6105704	43.854 ug/ml
Target Compounds			
2) N-DECANE	4.306	8116321	52.765 ug/ml
3) N-DODECANE	6.327	8125662	52.066 ug/ml
4) N-TETRADECANE	8.119	7996080	49.871 ug/ml
5) N-HEXADECANE	9.719	7804478	47.918 ug/ml
6) N-OCTADECANE	11.162	7578723	46.000 ug/ml
7) N-EICOSANE	12.472	7271436	44.448 ug/ml
8) N-DOCOSANE	13.671	7065841	43.781 ug/ml
10) N-TETRACOSANE	14.778	6916375	43.713 ug/ml
11) N-HEXACOSANE	15.800	6886332	44.617 ug/ml
12) N-OCTACOSANE	16.751	7079581	46.849 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032719\
Data File : FF004000.D
Signal(s) : FID2B.ch
Acq On : 27 Mar 2019 13:06
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Mar 28 00:40:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030119.M
Quant Title :
QLast Update : Mon Mar 04 08:35:38 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

