

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_F\Data\FF032723\  
Data File : FF012363.D  
Signal(s) : FID2B.ch  
Acq On : 27 Mar 2023 13:34  
Operator : YP\AJ  
Sample : 01950-19 5X  
Misc :  
ALS Vial : 65 Sample Multiplier: 1

Instrument :  
FID\_F  
ClientSampleId :  
RR-DIES-WP

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 03/28/2023  
Supervised By :Ankita Jodhani 03/28/2023

Integration File: autoint1.e  
Quant Time: Mar 28 08:04:58 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_F\Method\FF031623.M  
Quant Title :  
QLast Update : Wed Mar 15 18:07:23 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1uL  
Signal Phase : Rxi-1ms  
Signal Info : 20mx0.18mmx0.18um

| Compound                      | R.T.   | Response | Conc Units   |
|-------------------------------|--------|----------|--------------|
| -----                         |        |          |              |
| System Monitoring Compounds   |        |          |              |
| 9) S TETRACOSANE-d50 (SURR... | 15.024 | 345041   | 3.083 ug/mlm |
| Target Compounds              |        |          |              |
| -----                         |        |          |              |

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_F\Data\FF032723\  
Data File : FF012363.D  
Signal(s) : FID2B.ch  
Acq On : 27 Mar 2023 13:34  
Operator : YP\AJ  
Sample : 01950-19 5X  
Misc :  
ALS Vial : 65 Sample Multiplier: 1

Instrument :  
FID\_F  
ClientSampleId :  
RR-DIES-WP

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 03/28/2023  
Supervised By :Ankita Jodhani 03/28/2023

Integration File: autoint1.e  
Quant Time: Mar 28 08:04:58 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_F\Method\FF031623.M  
Quant Title :  
QLast Update : Wed Mar 15 18:07:23 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1uL  
Signal Phase : Rxi-1ms  
Signal Info : 20mx0.18mmx0.18um

