

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032922\
 Data File : FF010202.D
 Signal(s) : FID2B.ch
 Acq On : 29 Mar 2022 12:11
 Operator : AJ\MA
 Sample : N1645-04
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 30 01:56:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF032622.M
 Quant Title :
 QLast Update : Sat Mar 26 09:58:43 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.634	1905163	15.598 ug/mlm
Target Compounds			
1) N-OCTANE	2.280	323299	2.400 ug/ml
2) N-DECANE	4.368	354506	2.580 ug/ml
3) N-DODECANE	6.385	340249	2.461 ug/ml
4) N-TETRADECANE	8.180	321814	2.297 ug/mlm
5) N-HEXADECANE	9.781	328211	2.329 ug/ml
6) N-OCTADECANE	11.224	338378	2.348 ug/ml
7) N-EICOSANE	12.534	327618	2.343 ug/ml
8) N-DOCOSANE	13.733	313763	2.277 ug/ml
10) N-TETRACOSANE	14.837	308661	2.257 ug/ml
11) N-HEXACOSANE	15.859	291281	2.170 ug/ml
12) N-OCTACOSANE	16.810	289048	2.197 ug/ml
13) N-TRIACONTANE	17.697	296697	2.315 ug/ml
14) N-DOTRIACONTANE	18.531	286117	2.289 ug/ml
15) N-TETRATRIACONTANE	19.314	242032	2.241 ug/ml
16) N-HEXATRIACONTANE	20.053	205508	2.286 ug/mlm
17) N-OCTATRIACONTANE	20.781	168329	2.097 ug/ml
18) N-TETRACONTANE	21.670	174185	2.000 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032922\
Data File : FF010202.D
Signal(s) : FID2B.ch
Acq On : 29 Mar 2022 12:11
Operator : AJ\MA
Sample : N1645-04
Misc :
ALS Vial : 57 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 30 01:56:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF032622.M
Quant Title :
QLast Update : Sat Mar 26 09:58:43 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

