

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032922\
 Data File : FF010203.D
 Signal(s) : FID2B.ch
 Acq On : 29 Mar 2022 12:41
 Operator : AJ\MA
 Sample : N1645-05
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 30 01:56:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF032622.M
 Quant Title :
 QLast Update : Sat Mar 26 09:58:43 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.634	2090546	17.115 ug/mlm
Target Compounds			
1) N-OCTANE	2.280	663654	4.926 ug/ml
2) N-DECANE	4.368	745239	5.423 ug/ml
3) N-DODECANE	6.385	735148	5.317 ug/ml
4) N-TETRADECANE	8.180	700849	5.002 ug/ml
5) N-HEXADECANE	9.781	709235	5.033 ug/ml
6) N-OCTADECANE	11.224	718450	4.985 ug/ml
7) N-EICOSANE	12.535	695299	4.973 ug/ml
8) N-DOCOSANE	13.733	665167	4.827 ug/ml
10) N-TETRACOSANE	14.837	649464	4.748 ug/ml
11) N-HEXACOSANE	15.858	617456	4.599 ug/ml
12) N-OCTACOSANE	16.809	602423	4.579 ug/ml
13) N-TRIACONTANE	17.697	597718	4.665 ug/mlm
14) N-DOTRIACONTANE	18.530	565895	4.527 ug/ml
15) N-TETRATRIACONTANE	19.314	489345	4.530 ug/ml
16) N-HEXATRIACONTANE	20.053	402095	4.472 ug/ml
17) N-OCTATRIACONTANE	20.782	355049	4.423 ug/ml
18) N-TETRACONTANE	21.670	335576	3.853 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF032922\
Data File : FF010203.D
Signal(s) : FID2B.ch
Acq On : 29 Mar 2022 12:41
Operator : AJ\MA
Sample : N1645-05
Misc :
ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 30 01:56:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF032622.M
Quant Title :
QLast Update : Sat Mar 26 09:58:43 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

