

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF033022\
 Data File : FF010212.D
 Signal(s) : FID2B.ch
 Acq On : 30 Mar 2022 10:26
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 31 01:45:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF032622.M
 Quant Title :
 QLast Update : Sat Mar 26 09:58:43 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.634	6043735	49.481 ug/mlm
Target Compounds			
1) N-OCTANE	2.276	7246331	53.787 ug/ml
2) N-DECANE	4.369	7599208	55.300 ug/ml
3) N-DODECANE	6.387	7575220	54.788 ug/ml
4) N-TETRADECANE	8.183	7542649	53.828 ug/ml
5) N-HEXADECANE	9.784	7504180	53.251 ug/ml
6) N-OCTADECANE	11.226	7579659	52.590 ug/ml
7) N-EICOSANE	12.536	7204970	51.529 ug/ml
8) N-DOCOSANE	13.735	6957085	50.487 ug/ml
10) N-TETRACOSANE	14.840	6753026	49.374 ug/ml
11) N-HEXACOSANE	15.860	6507167	48.468 ug/ml
12) N-OCTACOSANE	16.811	6326118	48.090 ug/ml
13) N-TRIACONTANE	17.698	6167566	48.133 ug/ml
14) N-DOTRIACONTANE	18.531	5919911	47.360 ug/mlm
15) N-TETRATRIACONTANE	19.314	5386517	49.868 ug/mlm
16) N-HEXATRIACONTANE	20.054	4635755	51.563 ug/ml
17) N-OCTATRIACONTANE	20.780	4126211	51.404 ug/ml
18) N-TETRACONTANE	21.668	3765305	43.235 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF033022\
Data File : FF010212.D
Signal(s) : FID2B.ch
Acq On : 30 Mar 2022 10:26
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 31 01:45:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF032622.M
Quant Title :
QLast Update : Sat Mar 26 09:58:43 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

