

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF040422\
 Data File : FF010251.D
 Signal(s) : FID2B.ch
 Acq On : 04 Apr 2022 13:46
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/05/2022
 Supervised By :mohammad ahmed 04/06/2022

Integration File: autoint1.e
 Quant Time: Apr 05 01:12:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040122.M
 Quant Title :
 QLast Update : Fri Apr 01 11:58:56 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.625	5853896	48.975 ug/mlm
Target Compounds			
2) N-DECANE	4.364	6710612	50.475 ug/ml
3) N-DODECANE	6.381	6709517	50.391 ug/ml
4) N-TETRADECANE	8.175	6703486	50.559 ug/ml
5) N-HEXADECANE	9.776	6709971	50.542 ug/ml
6) N-OCTADECANE	11.217	6867908	49.952 ug/ml
7) N-EICOSANE	12.528	6662952	49.593 ug/ml
8) N-DOCOSANE	13.726	6612658	49.422 ug/ml
10) N-TETRACOSANE	14.830	6595253	49.252 ug/ml
11) N-HEXACOSANE	15.851	6516354	49.129 ug/ml
12) N-OCTACOSANE	16.801	6480674	48.896 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF040422\
Data File : FF010251.D
Signal(s) : FID2B.ch
Acq On : 04 Apr 2022 13:46
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/05/2022
Supervised By :mohammad ahmed 04/06/2022

Integration File: autoint1.e
Quant Time: Apr 05 01:12:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040122.M
Quant Title :
QLast Update : Fri Apr 01 11:58:56 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

