

rteres

Instrument :

FID_F

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/05/2022

Supervised By :mohammad ahmed 04/06/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF04042
Data File : FF010251.D
Signal(s) : FID2B.ch
Acq On : 04 Apr 2022 13:46
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040122.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.364	4.332	4.425	BB	756018	6710612	97.71%	9.264%
2	6.381	6.342	6.438	BB	692293	6709517	97.69%	9.262%
3	8.175	8.135	8.238	BB	642308	6703486	97.61%	9.254%
4	9.776	9.735	9.838	BB	617521	6709971	97.70%	9.263%
5	11.218	11.180	11.285	VB	593275	6867908	100.00%	9.481%
6	12.528	12.485	12.592	BB	555468	6662952	97.02%	9.198%
7	13.726	13.655	13.802	BB	531212	6612658	96.28%	9.129%
8	14.626	14.465	14.698	BB	426863	5868904	85.45%	8.102%
9	14.830	14.762	14.918	BB	512955	6595253	96.03%	9.105%
10	15.851	15.815	15.928	BB	491303	6516354	94.88%	8.996%
11	16.801	16.742	16.885	BB	479499	6480674	94.36%	8.946%
Sum of corrected areas:						72438289		

FF040122.M Tue Apr 05 21:21:49 2022