

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF040621\
 Data File : FF008237.D
 Signal(s) : FID2B.ch
 Acq On : 06 Apr 2021 08:58
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 07 01:40:30 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040521.M
 Quant Title :
 QLast Update : Tue Apr 06 02:37:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.794	5550432	50.580 ug/ml
Target Compounds			
1) N-OCTANE	2.443	6114908	50.354 ug/ml
2) N-DECANE	4.553	6361637	50.490 ug/ml
3) N-DODECANE	6.567	6435890	50.679 ug/ml
4) N-TETRADECANE	8.355	6572933	50.672 ug/ml
5) N-HEXADECANE	9.949	6505668	50.643 ug/ml
6) N-OCTADECANE	11.386	6531833	50.683 ug/ml
7) N-EICOSANE	12.693	6426094	50.496 ug/ml
8) N-DOCOSANE	13.893	6333372	50.461 ug/ml
10) N-TETRACOSANE	14.999	6274687	50.350 ug/ml
11) N-HEXACOSANE	16.019	6150909	50.252 ug/ml
12) N-OCTACOSANE	16.970	6099328	49.957 ug/ml
13) N-TRIACONTANE	17.852	6004958	49.139 ug/ml
14) N-DOTRIACONTANE	18.683	5836425	48.124 ug/ml
15) N-TETRATRIACONTANE	19.461	5578973	49.316 ug/ml
16) N-HEXATRIACONTANE	20.197	5146510	50.010 ug/ml
17) N-OCTATRIACONTANE	20.953	4734031	50.397 ug/ml
18) N-TETRACONTANE	21.902	4590707	50.429 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF040621\
Data File : FF008237.D
Signal(s) : FID2B.ch
Acq On : 06 Apr 2021 08:58
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 07 01:40:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040521.M
Quant Title :
QLast Update : Tue Apr 06 02:37:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

