

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF040721\
 Data File : FF008251.D
 Signal(s) : FID2B.ch
 Acq On : 07 Apr 2021 10:27
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 08 01:36:51 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040521.M
 Quant Title :
 QLast Update : Tue Apr 06 02:37:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.790	5715690	52.086 ug/ml
Target Compounds			
1) N-OCTANE	2.441	6428546	52.936 ug/ml
2) N-DECANE	4.549	6752884	53.595 ug/ml
3) N-DODECANE	6.563	6805920	53.593 ug/ml
4) N-TETRADECANE	8.351	6914572	53.306 ug/ml
5) N-HEXADECANE	9.945	6807560	52.993 ug/ml
6) N-OCTADECANE	11.382	6809819	52.840 ug/ml
7) N-EICOSANE	12.689	6681368	52.502 ug/ml
8) N-DOCOSANE	13.889	6559173	52.260 ug/ml
10) N-TETRACOSANE	14.995	6470252	51.920 ug/ml
11) N-HEXACOSANE	16.016	6318571	51.622 ug/ml
12) N-OCTACOSANE	16.965	6249579	51.188 ug/ml
13) N-TRIACONTANE	17.850	6151066	50.335 ug/ml
14) N-DOTRIACONTANE	18.678	5991781	49.405 ug/ml
15) N-TETRATRIACONTANE	19.457	5738771	50.729 ug/ml
16) N-HEXATRIACONTANE	20.192	5374845	52.228 ug/ml
17) N-OCTATRIACONTANE	20.945	4999896	53.227 ug/mlm
18) N-TETRACONTANE	21.892	4798694	52.714 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF040721\
Data File : FF008251.D
Signal(s) : FID2B.ch
Acq On : 07 Apr 2021 10:27
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 08 01:36:51 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040521.M
Quant Title :
QLast Update : Tue Apr 06 02:37:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

