

rteres

Instrument :

FID_F

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/11/2022

Supervised By :mohammad ahmed 04/12/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF04092
Data File : FF010311.D
Signal(s) : FID2B.ch
Acq On : 09 Apr 2022 09:13
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040122.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.272	2.245	2.318	BB	894170	6667518	97.04%	6.022%
2	4.365	4.328	4.421	BB	757421	6844846	99.62%	6.182%
3	6.384	6.341	6.448	BB	694904	6760041	98.38%	6.105%
4	8.179	8.141	8.248	BB	647795	6735576	98.03%	6.083%
5	9.780	9.738	9.851	BB	607117	6737305	98.05%	6.085%
6	11.223	11.155	11.291	BB	596648	6871129	100.00%	6.206%
7	12.533	12.491	12.608	BB	554704	6637578	96.60%	5.995%
8	13.732	13.695	13.798	VB	522258	6555918	95.41%	5.921%
9	14.632	14.468	14.705	BB	420637	5840761	85.00%	5.275%
10	14.836	14.768	14.905	BB	508605	6528722	95.02%	5.896%
11	15.857	15.815	15.921	BB	483260	6440149	93.73%	5.816%
12	16.807	16.741	16.868	BB	473915	6384595	92.92%	5.766%
13	17.694	17.621	17.745	BB	448350	6300542	91.70%	5.690%
14	18.527	18.435	18.575	BB	427231	6074057	88.40%	5.486%
15	19.310	19.261	19.349	BV	386038	5578162	81.18%	5.038%
16	20.050	19.993	20.103	M	331720	4940574	71.90%	4.462%
17	20.775	20.711	20.841	BB	254821	4447611	64.73%	4.017%
18	21.661	21.558	21.738	BB	192484	4380234	63.75%	3.956%
Sum of corrected areas:						110725319		

FF040122.M Mon Apr 11 11:37:50 2022