

rteres.txt

Instrument :
FID_F
ClientSampleId :
MDL-WATER-03-QT2-2023

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/12/2023
Supervised By :Ankita Jodhani 04/12/2023

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF04112
Data File : FF012448.D
Signal(s) : FID2B.ch
Acq On : 11 Apr 2023 10:57
Sample : 02213-09
Misc : 1 PPM WATER MDL
ALS Vial : 63 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031623.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.580	4.546	4.621	BB	7213	83878	4.57%	2.585%
2	6.755	6.711	7.231	M	1359	125108	6.82%	3.855%
3	8.596	8.557	8.910	M	2204	120640	6.58%	3.717%
4	10.185	10.152	10.236	BB	4711	85347	4.65%	2.630%
5	11.617	11.589	11.666	BB	7117	109140	5.95%	3.363%
6	12.919	12.889	12.966	BB	9108	123970	6.76%	3.820%
7	14.112	14.092	14.153	VB	11611	152589	8.32%	4.702%
8	15.012	14.963	15.041	BV	128691	1834414	100.00%	56.524%
9	15.216	15.148	15.267	BB	13458	225262	12.28%	6.941%
10	16.241	16.198	16.293	BB	12223	186327	10.16%	5.741%
11	17.189	17.148	17.229	BB	13383	198702	10.83%	6.123%
Sum of corrected areas:						3245378		

FF031623.M Wed Apr 12 03:28:25 2023