

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041321\
 Data File : FF008261.D
 Signal(s) : FID2B.ch
 Acq On : 13 Apr 2021 10:15
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 14 01:35:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040521.M
 Quant Title :
 QLast Update : Tue Apr 06 02:37:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.783	5265748	47.986 ug/ml
Target Compounds				
2)	N-DECANE	4.531	6477123	51.406 ug/ml
3)	N-DODECANE	6.549	6519038	51.334 ug/ml
4)	N-TETRADECANE	8.340	6581168	50.736 ug/ml
5)	N-HEXADECANE	9.936	6427662	50.035 ug/ml
6)	N-OCTADECANE	11.374	6372561	49.447 ug/ml
7)	N-EICOSANE	12.684	6196609	48.693 ug/ml
8)	N-DOCOSANE	13.883	6035787	48.090 ug/ml
10)	N-TETRACOSANE	14.988	5927712	47.566 ug/ml
11)	N-HEXACOSANE	16.009	5761382	47.070 ug/ml
12)	N-OCTACOSANE	16.956	5690314	46.607 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041321\
Data File : FF008261.D
Signal(s) : FID2B.ch
Acq On : 13 Apr 2021 10:15
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 14 01:35:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040521.M
Quant Title :
QLast Update : Tue Apr 06 02:37:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

