

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041522\
 Data File : FF010352.D
 Signal(s) : FID2B.ch
 Acq On : 15 Apr 2022 11:54
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 15 21:23:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040122.M
 Quant Title :
 QLast Update : Fri Apr 01 11:58:56 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.619	5516842	46.155 ug/ml
Target Compounds			
2) N-DECANE	4.359	6579212	49.487 ug/ml
3) N-DODECANE	6.376	6490053	48.743 ug/ml
4) N-TETRADECANE	8.170	6440843	48.578 ug/ml
5) N-HEXADECANE	9.771	6411140	48.291 ug/ml
6) N-OCTADECANE	11.212	6535421	47.534 ug/ml
7) N-EICOSANE	12.521	6310375	46.969 ug/ml
8) N-DOCOSANE	13.720	6235253	46.601 ug/ml
10) N-TETRACOSANE	14.823	6208827	46.367 ug/ml
11) N-HEXACOSANE	15.843	6133088	46.239 ug/mlm
12) N-OCTACOSANE	16.792	6061343	45.732 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041522\
Data File : FF010352.D
Signal(s) : FID2B.ch
Acq On : 15 Apr 2022 11:54
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 15 21:23:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040122.M
Quant Title :
QLast Update : Fri Apr 01 11:58:56 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

