

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041719\
 Data File : FF004238.D
 Signal(s) : FID2B.ch
 Acq On : 17 Apr 2019 10:30
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 18 00:46:24 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040219.M
 Quant Title :
 QLast Update : Tue Apr 02 13:45:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.537	6330369	48.695 ug/ml
Target Compounds			
1) N-OCTANE	2.197	7598266	54.262 ug/ml
2) N-DECANE	4.284	7736823	57.169 ug/ml
3) N-DODECANE	6.303	7788670	55.515 ug/ml
4) N-TETRADECANE	8.093	7759095	54.569 ug/ml
5) N-HEXADECANE	9.692	7693625	51.982 ug/ml
6) N-OCTADECANE	11.132	7621722	50.610 ug/ml
7) N-EICOSANE	12.440	7467575	49.691 ug/ml
8) N-DOCOSANE	13.638	7331017	48.978 ug/ml
10) N-TETRACOSANE	14.742	7162050	48.623 ug/ml
11) N-HEXACOSANE	15.762	7031386	48.435 ug/ml
12) N-OCTACOSANE	16.713	6936773	48.243 ug/ml
13) N-TRIACONTANE	17.600	6874082	48.132 ug/ml
14) N-DOTRIACONTANE	18.433	6627442	46.948 ug/ml
15) N-TETRATRIACONTANE	19.219	6490736	48.669 ug/ml
16) N-HEXATRIACONTANE	19.960	6260214	49.338 ug/ml
17) N-OCTATRIACONTANE	20.682	5744232	50.462 ug/ml
18) N-TETRACONTANE	21.544	5746459	47.376 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041719\
Data File : FF004238.D
Signal(s) : FID2B.ch
Acq On : 17 Apr 2019 10:30
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 18 00:46:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040219.M
Quant Title :
QLast Update : Tue Apr 02 13:45:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

