

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041719\
 Data File : FF004240.D
 Signal(s) : FID2B.ch
 Acq On : 17 Apr 2019 13:06
 Operator : AJ\SJ
 Sample : PB118959BS
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 18 00:46:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040219.M
 Quant Title :
 QLast Update : Tue Apr 02 13:45:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.536	2228905	17.145 ug/ml
Target Compounds			
1) N-OCTANE	2.198	2628461	18.771 ug/ml
2) N-DECANE	4.284	2592896	19.159 ug/ml
3) N-DODECANE	6.302	2637177	18.797 ug/ml
4) N-TETRADECANE	8.092	2638242	18.554 ug/ml
5) N-HEXADECANE	9.690	2662337	17.988 ug/ml
6) N-OCTADECANE	11.130	2665098	17.697 ug/ml
7) N-EICOSANE	12.439	2608114	17.355 ug/ml
8) N-DOCOSANE	13.636	2565882	17.142 ug/ml
10) N-TETRACOSANE	14.740	2515893	17.080 ug/ml
11) N-HEXACOSANE	15.761	2481431	17.093 ug/ml
12) N-OCTACOSANE	16.711	2489967	17.317 ug/ml
13) N-TRIACONTANE	17.599	2522236	17.661 ug/ml
14) N-DOTRIACONTANE	18.431	2478746	17.559 ug/ml
15) N-TETRATRIACONTANE	19.216	2439778	18.294 ug/ml
16) N-HEXATRIACONTANE	19.960	2378645	18.746 ug/ml
17) N-OCTATRIACONTANE	20.680	2188337	19.224 ug/ml
18) N-TETRACONTANE	21.543	2190293	18.057 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041719\
Data File : FF004240.D
Signal(s) : FID2B.ch
Acq On : 17 Apr 2019 13:06
Operator : AJ\SJ
Sample : PB118959BS
Misc :
ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 18 00:46:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040219.M
Quant Title :
QLast Update : Tue Apr 02 13:45:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

