

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041919\
 Data File : FF004277.D
 Signal(s) : FID2B.ch
 Acq On : 19 Apr 2019 18:14
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 19 23:20:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040219.M
 Quant Title :
 QLast Update : Tue Apr 02 13:45:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.538	5843979	44.954 ug/ml
Target Compounds			
2) N-DECANE	4.284	7437962	54.961 ug/ml
3) N-DODECANE	6.302	7505099	53.494 ug/ml
4) N-TETRADECANE	8.093	7453916	52.423 ug/ml
5) N-HEXADECANE	9.690	7342660	49.611 ug/ml
6) N-OCTADECANE	11.131	7220545	47.946 ug/ml
7) N-EICOSANE	12.440	6992790	46.532 ug/ml
8) N-DOCOSANE	13.638	6798878	45.423 ug/ml
10) N-TETRACOSANE	14.742	6620429	44.946 ug/ml
11) N-HEXACOSANE	15.764	6523573	44.937 ug/ml
12) N-OCTACOSANE	16.715	6514860	45.308 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF041919\
Data File : FF004277.D
Signal(s) : FID2B.ch
Acq On : 19 Apr 2019 18:14
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 19 23:20:59 2019
Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040219.M
Quant Title :
QLast Update : Tue Apr 02 13:45:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

