

rters

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/25/2022
Supervised By :mohammad ahmed 04/25/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF04222
Data File : FF010373.D
Signal(s) : FID2B.ch
Acq On : 22 Apr 2022 13:21
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040122.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.267	2.238	2.321	BB	887791	6751551	100.00%	6.450%
2	4.358	4.325	4.428	BB	735824	6684249	99.00%	6.386%
3	6.374	6.335	6.445	BB	664145	6554586	97.08%	6.262%
4	8.168	8.131	8.238	BB	624997	6485350	96.06%	6.196%
5	9.769	9.725	9.831	BB	584054	6443271	95.43%	6.155%
6	11.209	11.141	11.285	BB	552445	6560341	97.17%	6.267%
7	12.518	12.475	12.595	BB	524693	6298376	93.29%	6.017%
8	13.715	13.641	13.801	BB	497139	6171921	91.41%	5.896%
9	14.614	14.568	14.708	BB	390278	5419718	80.27%	5.178%
10	14.818	14.748	14.905	BB	468100	6086344	90.15%	5.814%
11	15.837	15.794	15.880	M	453713	5941000	87.99%	5.676%
12	16.787	16.741	16.855	BB	434788	5824589	86.27%	5.564%
13	17.673	17.605	17.738	BB	409432	5714888	84.65%	5.459%
14	18.504	18.428	18.581	BB	382755	5594714	82.87%	5.345%
15	19.285	19.244	19.345	M	358188	5115301	75.76%	4.887%
16	20.022	19.979	20.056	M	319067	4604277	68.20%	4.399%
17	20.745	20.694	20.813	M	246171	4178526	61.89%	3.992%
18	21.620	21.552	21.717	M	177664	4248918	62.93%	4.059%
Sum of corrected areas:						104677920		

FF040122.M Mon Apr 25 18:55:31 2022