

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF042619\
 Data File : FF004374.D
 Signal(s) : FID2B.ch
 Acq On : 26 Apr 2019 21:20
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Apr 26 22:56:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040219.M
 Quant Title :
 QLast Update : Tue Apr 02 13:45:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.523	5584370	42.957 ug/ml
Target Compounds			
1) N-OCTANE	2.190	6702238	47.863 ug/mlm
2) N-DECANE	4.275	6870259	50.766 ug/ml
3) N-DODECANE	6.291	7008362	49.953 ug/ml
4) N-TETRADECANE	8.080	7003138	49.252 ug/ml
5) N-HEXADECANE	9.677	6961822	47.038 ug/ml
6) N-OCTADECANE	11.117	6857090	45.532 ug/ml
7) N-EICOSANE	12.426	6659095	44.311 ug/ml
8) N-DOCOSANE	13.624	6474680	43.257 ug/ml
10) N-TETRACOSANE	14.727	6341411	43.051 ug/ml
11) N-HEXACOSANE	15.749	6282501	43.276 ug/ml
12) N-OCTACOSANE	16.700	6327810	44.008 ug/ml
13) N-TRIACONTANE	17.590	6453571	45.188 ug/ml
14) N-DOTRIACONTANE	18.426	6409500	45.404 ug/ml
15) N-TETRATRIACONTANE	19.213	6350350	47.617 ug/ml
16) N-HEXATRIACONTANE	19.957	6310547	49.734 ug/ml
17) N-OCTATRIACONTANE	20.681	5812913	51.065 ug/mlm
18) N-TETRACONTANE	21.542	5998521	49.454 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF042619\
Data File : FF004374.D
Signal(s) : FID2B.ch
Acq On : 26 Apr 2019 21:20
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Apr 26 22:56:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040219.M
Quant Title :
QLast Update : Tue Apr 02 13:45:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

