

rters

Instrument :

FID_F

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/28/2022

Supervised By :mohammad ahmed 04/29/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF04272
Data File : FF010411.D
Signal(s) : FID2B.ch
Acq On : 27 Apr 2022 12:36
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF040122.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.263	2.228	2.315	BB	879213	6674432	95.14%	5.931%
2	4.354	4.315	4.428	BB	768029	6964046	99.26%	6.188%
3	6.372	6.325	6.442	BB	696062	6910534	98.50%	6.141%
4	8.166	8.125	8.235	BB	654094	6896756	98.31%	6.129%
5	9.767	9.725	9.835	BB	627621	6884457	98.13%	6.118%
6	11.208	11.138	11.275	BB	623237	7015621	100.00%	6.234%
7	12.518	12.475	12.578	BB	576986	6762576	96.39%	6.009%
8	13.715	13.638	13.785	BB	539249	6655995	94.87%	5.915%
9	14.615	14.568	14.722	BB	429241	5886432	83.90%	5.231%
10	14.818	14.752	14.878	BB	517181	6579523	93.78%	5.847%
11	15.839	15.797	15.879	M	500578	6449592	91.93%	5.731%
12	16.787	16.742	16.852	BB	468323	6356004	90.60%	5.648%
13	17.674	17.605	17.748	BB	458348	6303959	89.86%	5.602%
14	18.504	18.455	18.585	BB	427945	6137649	87.49%	5.454%
15	19.287	19.229	19.348	M	395764	5664316	80.74%	5.033%
16	20.025	19.825	20.122	BB	337630	5084929	72.48%	4.519%
17	20.745	20.696	20.792	M	268448	4653651	66.33%	4.135%
18	21.623	21.515	21.752	BB	199109	4652723	66.32%	4.135%
Sum of corrected areas:						112533195		

FF040122.M Thu Apr 28 10:55:07 2022