

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050123\
 Data File : FF012554.D
 Signal(s) : FID2B.ch
 Acq On : 01 May 2023 15:04
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/02/2023
 Supervised By :Ankita Jodhani 05/02/2023

Integration File: autoint1.e
 Quant Time: May 02 01:14:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF041323.M
 Quant Title :
 QLast Update : Thu Apr 13 15:26:28 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.012	5094101	42.379 ug/ml
Target Compounds			
1) N-OCTANE	2.056	5055587	40.624 ug/ml
2) N-DECANE	4.573	5775338	46.349 ug/ml
3) N-DODECANE	6.736	5854827	47.405 ug/ml
4) N-TETRADECANE	8.561	5857432	45.663 ug/ml
5) N-HEXADECANE	10.166	6033007	45.184 ug/ml
6) N-OCTADECANE	11.608	6084185	44.601 ug/ml
7) N-EICOSANE	12.917	6039468	43.858 ug/ml
8) N-DOCOSANE	14.114	5817745	43.219 ug/ml
10) N-TETRACOSANE	15.216	5768053	42.642 ug/ml
11) N-HEXACOSANE	16.236	5591717	41.993 ug/ml
12) N-OCTACOSANE	17.185	5505881	41.747 ug/ml
13) N-TRIACONTANE	18.072	5501757	42.163 ug/ml
14) N-DOTRIACONTANE	18.905	5307121	42.070 ug/ml
15) N-TETRATRIACONTANE	19.687	4881950	43.049 ug/ml
16) N-HEXATRIACONTANE	20.427	4459620	43.673 ug/mlm
17) N-OCTATRIACONTANE	21.161	4632456	47.537 ug/ml
18) N-TETRACONTANE	22.066	4471110	46.233 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050123\
Data File : FF012554.D
Signal(s) : FID2B.ch
Acq On : 01 May 2023 15:04
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/02/2023
Supervised By :Ankita Jodhani 05/02/2023

Integration File: autoint1.e
Quant Time: May 02 01:14:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF041323.M
Quant Title :
QLast Update : Thu Apr 13 15:26:28 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

