

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050523\
 Data File : FF012582.D
 Signal(s) : FID2B.ch
 Acq On : 05 May 2023 13:47
 Operator : YP\AJ
 Sample : 100 TRPH STD
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 05 16:12:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050523.M
 Quant Title :
 QLast Update : Fri May 05 16:11:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.006	11071859	89.197 ug/ml
Target Compounds			
1) N-OCTANE	2.054	12290450	91.078 ug/ml
2) N-DECANE	4.570	12736350	90.331 ug/ml
3) N-DODECANE	6.734	12928613	90.575 ug/ml
4) N-TETRADECANE	8.559	12937128	90.447 ug/ml
5) N-HEXADECANE	10.165	13176063	90.058 ug/ml
6) N-OCTADECANE	11.606	13185893	89.877 ug/ml
7) N-EICOSANE	12.913	13063187	89.813 ug/ml
8) N-DOCOSANE	14.109	12559284	89.945 ug/ml
10) N-TETRACOSANE	15.211	12424124	89.857 ug/ml
11) N-HEXACOSANE	16.230	12066956	89.753 ug/ml
12) N-OCTACOSANE	17.178	11706891	89.294 ug/ml
13) N-TRIACONTANE	18.065	11341180	88.079 ug/ml
14) N-DOTRIACONTANE	18.895	10680851	87.272 ug/ml
15) N-TETRATRIACONTANE	19.678	9905959	87.702 ug/ml
16) N-HEXATRIACONTANE	20.416	9425887	87.078 ug/ml
17) N-OCTATRIACONTANE	21.147	9193919	88.791 ug/ml
18) N-TETRACONTANE	22.050	9512171	89.213 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050523\
Data File : FF012582.D
Signal(s) : FID2B.ch
Acq On : 05 May 2023 13:47
Operator : YP\AJ
Sample : 100 TRPH STD
Misc :
ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 05 16:12:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050523.M
Quant Title :
QLast Update : Fri May 05 16:11:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

