

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050622\
 Data File : FF010571.D
 Signal(s) : FID2B.ch
 Acq On : 06 May 2022 15:08
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: May 08 22:53:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF042922.M
 Quant Title :
 QLast Update : Fri Apr 29 04:08:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.610	5414015	44.146 ug/ml
Target Compounds			
2) N-DECANE	4.352	6628077	49.167 ug/ml
3) N-DODECANE	6.370	6610912	48.774 ug/ml
4) N-TETRADECANE	8.164	6604141	48.407 ug/ml
5) N-HEXADECANE	9.764	6577011	47.544 ug/ml
6) N-OCTADECANE	11.205	6658801	46.373 ug/ml
7) N-EICOSANE	12.514	6357095	45.387 ug/ml
8) N-DOCOSANE	13.712	6192071	44.582 ug/ml
10) N-TETRACOSANE	14.814	6091049	44.087 ug/ml
11) N-HEXACOSANE	15.833	5950556	43.974 ug/ml
12) N-OCTACOSANE	16.783	5800353	44.110 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050622\
Data File : FF010571.D
Signal(s) : FID2B.ch
Acq On : 06 May 2022 15:08
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: May 08 22:53:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF042922.M
Quant Title :
QLast Update : Fri Apr 29 04:08:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

