

rteres

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/09/2022
Supervised By :mohammad ahmed 05/09/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF05062
Data File : FF010569.D
Signal(s) : FID2B.ch
Acq On : 06 May 2022 13:00
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF042922.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.261	2.231	2.321	BB	871518	6442985	95.89%	6.266%
2	4.353	4.318	4.428	BB	734822	6563674	97.69%	6.383%
3	6.371	6.335	6.445	BB	666391	6563971	97.69%	6.384%
4	8.165	8.128	8.238	BB	618166	6584908	98.00%	6.404%
5	9.766	9.721	9.838	BB	594090	6597531	98.19%	6.416%
6	11.206	11.138	11.275	BB	576722	6718977	100.00%	6.534%
7	12.515	12.471	12.578	BB	537886	6458698	96.13%	6.281%
8	13.713	13.635	13.775	BB	509614	6326277	94.16%	6.152%
9	14.611	14.565	14.660	BBA	407501	5542189	82.49%	5.390%
10	14.815	14.745	14.878	BB	493730	6232993	92.77%	6.062%
11	15.835	15.781	15.898	BB	465375	6084256	90.55%	5.917%
12	16.783	16.741	16.830	BBA	434129	5907740	87.93%	5.745%
13	17.669	17.598	17.735	BB	396799	5699650	84.83%	5.543%
14	18.499	18.428	18.551	BB	370573	5242076	78.02%	5.098%
15	19.281	19.085	19.331	BB	325409	4579495	68.16%	4.454%
16	20.020	19.971	20.060	BBA	268970	3981852	59.26%	3.872%
17	20.741	20.645	20.828	BB	210919	3653696	54.38%	3.553%
18	21.616	21.511	21.698	BB	162307	3643802	54.23%	3.544%
Sum of corrected areas:						102824770		

FF042922.M Mon May 09 13:57:28 2022