

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050720\
 Data File : FF006871.D
 Signal(s) : FID2B.ch
 Acq On : 07 May 2020 20:35
 Operator : DD\AJ
 Sample : PB128707BS
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 PB128707BS

Integration File: Sample.e
 Quant Time: May 08 01:03:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF042820.M
 Quant Title :
 QLast Update : Wed Apr 29 02:00:23 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.575	2280474	16.787 ug/ml
Target Compounds			
2) N-DECANE	4.345	2802749	20.350 ug/ml
3) N-DODECANE	6.358	2844403	19.949 ug/ml
4) N-TETRADECANE	8.148	2801376	19.289 ug/ml
5) N-HEXADECANE	9.744	2787646	18.923 ug/ml
6) N-OCTADECANE	11.181	2853271	18.319 ug/ml
7) N-EICOSANE	12.486	2742760	17.746 ug/ml
8) N-DOCOSANE	13.680	2706580	17.280 ug/ml
10) N-TETRACOSANE	14.779	2646571	16.981 ug/ml
11) N-HEXACOSANE	15.797	2589928	16.761 ug/ml
12) N-OCTACOSANE	16.742	2556815	16.588 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050720\
Data File : FF006871.D
Signal(s) : FID2B.ch
Acq On : 07 May 2020 20:35
Operator : DD\AJ
Sample : PB128707BS
Misc :
ALS Vial : 72 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :
PB128707BS

Integration File: Sample.e
Quant Time: May 08 01:03:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF042820.M
Quant Title :
QLast Update : Wed Apr 29 02:00:23 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

