

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050819\
 Data File : FF004471.D
 Signal(s) : FID2B.ch
 Acq On : 09 May 2019 00:09
 Operator : AJ\SJ
 Sample : 100 TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 09 09:28:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050819.M
 Quant Title :
 QLast Update : Thu May 09 09:25:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.682	10155635	105.113 ug/ml
Target Compounds			
1) N-OCTANE	2.313	14396029	101.874 ug/ml
2) N-DECANE	4.417	14215674	102.306 ug/ml
3) N-DODECANE	6.438	13928615	102.509 ug/ml
4) N-TETRADECANE	8.231	13410685	102.821 ug/ml
5) N-HEXADECANE	9.831	13047246	103.416 ug/ml
6) N-OCTADECANE	11.273	12468740	103.951 ug/ml
7) N-EICOSANE	12.583	12093188	103.884 ug/ml
8) N-DOCOSANE	13.781	11650782	104.041 ug/ml
10) N-TETRACOSANE	14.886	11606584	105.193 ug/ml
11) N-HEXACOSANE	15.909	11697982	105.929 ug/ml
12) N-OCTACOSANE	16.858	11701056	105.144 ug/ml
13) N-TRIACONTANE	17.747	11688628	103.479 ug/ml
14) N-DOTRIACONTANE	18.582	11296643	99.116 ug/ml
15) N-TETRATRIACONTANE	19.369	11038210	97.764 ug/ml
16) N-HEXATRIACONTANE	20.111	10337684	96.907 ug/ml
17) N-OCTATRIACONTANE	20.858	9533766	94.069 ug/ml
18) N-TETRACONTANE	21.774	9409593	78.199 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050819\
Data File : FF004471.D
Signal(s) : FID2B.ch
Acq On : 09 May 2019 00:09
Operator : AJ\SJ
Sample : 100 TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 09 09:28:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050819.M
Quant Title :
QLast Update : Thu May 09 09:25:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

