

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050819\
 Data File : FF004475.D
 Signal(s) : FID2B.ch
 Acq On : 09 May 2019 02:06
 Operator : AJ\SJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 09 09:28:21 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050819.M
 Quant Title :
 QLast Update : Thu May 09 09:25:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.675	524428	5.428 ug/ml
Target Compounds			
1) N-OCTANE	2.315	776418	5.494 ug/ml
2) N-DECANE	4.414	766522	5.516 ug/ml
3) N-DODECANE	6.432	754151	5.550 ug/ml
4) N-TETRADECANE	8.225	726002	5.566 ug/ml
5) N-HEXADECANE	9.825	694243	5.503 ug/ml
6) N-OCTADECANE	11.266	653957	5.452 ug/ml
7) N-EICOSANE	12.577	634195	5.448 ug/ml
8) N-DOCOSANE	13.776	609798	5.445 ug/ml
10) N-TETRACOSANE	14.879	597635	5.416 ug/ml
11) N-HEXACOSANE	15.901	593099	5.371 ug/ml
12) N-OCTACOSANE	16.851	589599	5.298 ug/ml
13) N-TRIACONTANE	17.743	602158	5.331 ug/ml
14) N-DOTRIACONTANE	18.578	679836	5.965 ug/ml
15) N-TETRATRIACONTANE	19.365	609416	5.398 ug/ml
16) N-HEXATRIACONTANE	20.110	562990	5.278 ug/ml
17) N-OCTATRIACONTANE	20.854	523683	5.167 ug/ml
18) N-TETRACONTANE	21.771	631549	5.249 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050819\
Data File : FF004475.D
Signal(s) : FID2B.ch
Acq On : 09 May 2019 02:06
Operator : AJ\SJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 57 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 09 09:28:21 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050819.M
Quant Title :
QLast Update : Thu May 09 09:25:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

