

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050922\
 Data File : FF010582.D
 Signal(s) : FID2B.ch
 Acq On : 09 May 2022 09:57
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: Sample.e
 Quant Time: May 10 01:27:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF042922.M
 Quant Title :
 QLast Update : Fri Apr 29 04:08:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.611	6006755	48.979 ug/ml
Target Compounds			
2) N-DECANE	4.354	7156269	53.085 ug/ml
3) N-DODECANE	6.371	7128611	52.594 ug/ml
4) N-TETRADECANE	8.166	7140090	52.335 ug/ml
5) N-HEXADECANE	9.765	7148768	51.677 ug/ml
6) N-OCTADECANE	11.207	7274975	50.665 ug/ml
7) N-EICOSANE	12.515	6988730	49.896 ug/ml
8) N-DOCOSANE	13.713	6853001	49.340 ug/ml
10) N-TETRACOSANE	14.815	6795136	49.183 ug/ml
11) N-HEXACOSANE	15.836	6615554	48.888 ug/ml
12) N-OCTACOSANE	16.784	6447960	49.035 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF050922\
Data File : FF010582.D
Signal(s) : FID2B.ch
Acq On : 09 May 2022 09:57
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: Sample.e
Quant Time: May 10 01:27:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF042922.M
Quant Title :
QLast Update : Fri Apr 29 04:08:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

