

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051121\
 Data File : FF008391.D
 Signal(s) : FID2B.ch
 Acq On : 11 May 2021 09:30
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 12 01:23:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050521.M
 Quant Title :
 QLast Update : Thu May 06 08:43:31 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.781	6119631	49.014 ug/ml
Target Compounds			
1) N-OCTANE	2.429	6956740	52.478 ug/ml
2) N-DECANE	4.535	7245877	53.675 ug/ml
3) N-DODECANE	6.549	7326995	53.591 ug/ml
4) N-TETRADECANE	8.338	7427060	53.142 ug/ml
5) N-HEXADECANE	9.934	7322559	52.513 ug/ml
6) N-OCTADECANE	11.371	7393476	51.862 ug/ml
7) N-EICOSANE	12.680	7238206	50.952 ug/ml
8) N-DOCOSANE	13.880	7105710	49.924 ug/ml
10) N-TETRACOSANE	14.985	6939332	48.656 ug/ml
11) N-HEXACOSANE	16.007	6692780	47.468 ug/ml
12) N-OCTACOSANE	16.954	6502962	46.345 ug/ml
13) N-TRIACONTANE	17.837	6321186	45.620 ug/ml
14) N-DOTRIACONTANE	18.666	6039735	44.575 ug/ml
15) N-TETRATRIACONTANE	19.443	5764772	45.209 ug/ml
16) N-HEXATRIACONTANE	20.178	5466915	45.799 ug/ml
17) N-OCTATRIACONTANE	20.929	5137964	47.721 ug/ml
18) N-TETRACONTANE	21.867	5226868	45.976 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051121\
Data File : FF008391.D
Signal(s) : FID2B.ch
Acq On : 11 May 2021 09:30
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 12 01:23:15 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050521.M
Quant Title :
QLast Update : Thu May 06 08:43:31 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

