

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051421\
 Data File : FF008423.D
 Signal(s) : FID2B.ch
 Acq On : 14 May 2021 14:01
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 15 00:10:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050521.M
 Quant Title :
 QLast Update : Thu May 06 08:43:31 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.780	5503204	44.077 ug/ml
Target Compounds			
2) N-DECANE	4.536	6466000	47.898 ug/ml
3) N-DODECANE	6.549	6531248	47.771 ug/ml
4) N-TETRADECANE	8.339	6623056	47.389 ug/ml
5) N-HEXADECANE	9.934	6537754	46.885 ug/ml
6) N-OCTADECANE	11.371	6605310	46.333 ug/ml
7) N-EICOSANE	12.679	6483422	45.639 ug/ml
8) N-DOCOSANE	13.880	6378157	44.812 ug/ml
10) N-TETRACOSANE	14.985	6246185	43.796 ug/ml
11) N-HEXACOSANE	16.008	6054734	42.942 ug/ml
12) N-OCTACOSANE	16.954	5930671	42.266 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051421\
Data File : FF008423.D
Signal(s) : FID2B.ch
Acq On : 14 May 2021 14:01
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 15 00:10:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050521.M
Quant Title :
QLast Update : Thu May 06 08:43:31 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

