

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051623\
 Data File : FF012602.D
 Signal(s) : FID2B.ch
 Acq On : 16 May 2023 11:06
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 17 01:12:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050523.M
 Quant Title :
 QLast Update : Fri May 05 16:23:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.006	5779978	47.880 ug/ml
Target Compounds			
1) N-OCTANE	2.050	6214900	47.760 ug/ml
2) N-DECANE	4.565	6416098	48.064 ug/ml
3) N-DODECANE	6.729	6471640	49.522 ug/ml
4) N-TETRADECANE	8.554	6499922	50.434 ug/ml
5) N-HEXADECANE	10.160	6684897	49.600 ug/ml
6) N-OCTADECANE	11.602	6758689	48.793 ug/ml
7) N-EICOSANE	12.910	6712784	48.491 ug/ml
8) N-DOCOSANE	14.107	6455084	48.186 ug/ml
10) N-TETRACOSANE	15.210	6379972	47.813 ug/ml
11) N-HEXACOSANE	16.230	6172168	47.612 ug/ml
12) N-OCTACOSANE	17.178	5982273	47.206 ug/ml
13) N-TRIACONTANE	18.064	5823467	46.562 ug/ml
14) N-DOTRIACONTANE	18.896	5522922	46.233 ug/ml
15) N-TETRATRIACONTANE	19.678	5158409	46.893 ug/ml
16) N-HEXATRIACONTANE	20.418	4872904	46.830 ug/ml
17) N-OCTATRIACONTANE	21.152	4991190	49.632 ug/ml
18) N-TETRACONTANE	22.056	5294523	51.111 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051623\
Data File : FF012602.D
Signal(s) : FID2B.ch
Acq On : 16 May 2023 11:06
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 17 01:12:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050523.M
Quant Title :
QLast Update : Fri May 05 16:23:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

