

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051623\
 Data File : FF012611.D
 Signal(s) : FID2B.ch
 Acq On : 16 May 2023 15:58
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 17 01:13:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050523.M
 Quant Title :
 QLast Update : Fri May 05 16:23:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.008	5781841	47.896 ug/ml
Target Compounds			
1) N-OCTANE	2.053	6247469	48.010 ug/ml
2) N-DECANE	4.567	6485001	48.580 ug/ml
3) N-DODECANE	6.731	6556320	50.170 ug/ml
4) N-TETRADECANE	8.556	6556326	50.872 ug/ml
5) N-HEXADECANE	10.162	6747372	50.063 ug/ml
6) N-OCTADECANE	11.604	6801092	49.100 ug/ml
7) N-EICOSANE	12.912	6741141	48.696 ug/ml
8) N-DOCOSANE	14.110	6455257	48.187 ug/ml
10) N-TETRACOSANE	15.212	6373415	47.763 ug/ml
11) N-HEXACOSANE	16.232	6169466	47.591 ug/ml
12) N-OCTACOSANE	17.181	5990659	47.272 ug/ml
13) N-TRIACONTANE	18.068	5857171	46.832 ug/ml
14) N-DOTRIACONTANE	18.899	5592663	46.816 ug/ml
15) N-TETRATRIACONTANE	19.683	5230748	47.551 ug/ml
16) N-HEXATRIACONTANE	20.423	5011392	48.161 ug/ml
17) N-OCTATRIACONTANE	21.156	5177030	51.480 ug/ml
18) N-TETRACONTANE	22.061	5294909	51.114 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051623\
Data File : FF012611.D
Signal(s) : FID2B.ch
Acq On : 16 May 2023 15:58
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 17 01:13:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050523.M
Quant Title :
QLast Update : Fri May 05 16:23:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

