

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051719\
 Data File : FF004561.D
 Signal(s) : FID2B.ch
 Acq On : 17 May 2019 11:15
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: May 18 05:01:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050819.M
 Quant Title :
 QLast Update : Thu May 09 09:32:48 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.703	5003210	49.926 ug/ml
Target Compounds			
2) N-DECANE	4.423	6810949	47.180 ug/ml
3) N-DODECANE	6.446	6572676	46.468 ug/ml
4) N-TETRADECANE	8.241	6271201	46.149 ug/ml
5) N-HEXADECANE	9.844	6070931	46.378 ug/ml
6) N-OCTADECANE	11.288	5821502	46.962 ug/ml
7) N-EICOSANE	12.601	5748098	47.765 ug/ml
8) N-DOCOSANE	13.801	5684302	48.983 ug/ml
10) N-TETRACOSANE	14.908	5751502	50.233 ug/ml
11) N-HEXACOSANE	15.931	5821562	50.988 ug/ml
12) N-OCTACOSANE	16.885	5894845	51.583 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF051719\
Data File : FF004561.D
Signal(s) : FID2B.ch
Acq On : 17 May 2019 11:15
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: May 18 05:01:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF050819.M
Quant Title :
QLast Update : Thu May 09 09:32:48 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

