

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052020\
 Data File : FF006984.D
 Signal(s) : FID2B.ch
 Acq On : 20 May 2020 14:55
 Operator : DD\AJ
 Sample : L2182-09
 Misc : MDL-WATER-1 PPM
 ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 21 05:09:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF051520.M
 Quant Title :
 QLast Update : Sat May 16 03:36:49 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.723	2709635	19.605 ug/ml
Target Compounds			
2) N-DECANE	4.460	171716	1.042 ug/ml
3) N-DODECANE	6.479	176921	1.057 ug/ml
4) N-TETRADECANE	8.273	180880	1.081 ug/ml
5) N-HEXADECANE	9.873	179831	1.102 ug/ml
6) N-OCTADECANE	11.315	186687	1.111 ug/ml
7) N-EICOSANE	12.624	183018	1.121 ug/ml
8) N-DOCOSANE	13.821	184882	1.140 ug/ml
10) N-TETRACOSANE	14.924	181880	1.150 ug/ml
11) N-HEXACOSANE	15.945	182555	1.190 ug/ml
12) N-OCTACOSANE	16.895	200182	1.345 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052020\
Data File : FF006984.D
Signal(s) : FID2B.ch
Acq On : 20 May 2020 14:55
Operator : DD\AJ
Sample : L2182-09
Misc : MDL-WATER-1 PPM
ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 21 05:09:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF051520.M
Quant Title :
QLast Update : Sat May 16 03:36:49 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

