

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052219\
 Data File : FF004603.D
 Signal(s) : FID2B.ch
 Acq On : 22 May 2019 22:54
 Operator : AJ\SJ
 Sample : FF052219ICV
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 23 07:25:15 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
 Quant Title :
 QLast Update : Thu May 23 07:21:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.703	6218938	50.929 ug/ml
Target Compounds			
1) N-OCTANE	2.313	7508339	52.044 ug/ml
2) N-DECANE	4.420	7614113	51.998 ug/ml
3) N-DODECANE	6.444	7667274	51.812 ug/ml
4) N-TETRADECANE	8.239	7610940	51.814 ug/ml
5) N-HEXADECANE	9.842	7618101	51.597 ug/ml
6) N-OCTADECANE	11.286	7489769	51.588 ug/ml
7) N-EICOSANE	12.600	7447826	51.425 ug/ml
8) N-DOCOSANE	13.800	7243293	51.260 ug/ml
10) N-TETRACOSANE	14.908	7060291	50.945 ug/ml
11) N-HEXACOSANE	15.932	6923282	50.812 ug/ml
12) N-OCTACOSANE	16.886	6874383	50.562 ug/ml
13) N-TRIACONTANE	17.775	6914591	50.003 ug/ml
14) N-DOTRIACONTANE	18.612	7264899	52.517 ug/ml
15) N-TETRATRIACONTANE	19.400	6803644	49.763 ug/ml
16) N-HEXATRIACONTANE	20.146	6556104	49.799 ug/ml
17) N-OCTATRIACONTANE	20.898	6142410	51.079 ug/ml
18) N-TETRACONTANE	21.830	6363785	51.679 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052219\
Data File : FF004603.D
Signal(s) : FID2B.ch
Acq On : 22 May 2019 22:54
Operator : AJ\SJ
Sample : FF052219ICV
Misc :
ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 23 07:25:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
Quant Title :
QLast Update : Thu May 23 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

