

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052419\
 Data File : FF004628.D
 Signal(s) : FID2B.ch
 Acq On : 24 May 2019 19:04
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 24 22:50:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
 Quant Title :
 QLast Update : Thu May 23 07:21:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.703	5934097	48.596 ug/ml
Target Compounds			
1) N-OCTANE	2.312	7425627	51.471 ug/ml
2) N-DECANE	4.419	7513991	51.314 ug/ml
3) N-DODECANE	6.442	7495244	50.650 ug/ml
4) N-TETRADECANE	8.238	7392897	50.330 ug/ml
5) N-HEXADECANE	9.841	7363409	49.872 ug/ml
6) N-OCTADECANE	11.285	7200515	49.596 ug/ml
7) N-EICOSANE	12.597	7133099	49.252 ug/ml
8) N-DOCOSANE	13.800	6914174	48.931 ug/ml
10) N-TETRACOSANE	14.906	6735599	48.602 ug/ml
11) N-HEXACOSANE	15.931	6612176	48.529 ug/ml
12) N-OCTACOSANE	16.883	6585888	48.440 ug/ml
13) N-TRIACONTANE	17.775	6652001	48.104 ug/ml
14) N-DOTRIACONTANE	18.612	6658546	48.134 ug/ml
15) N-TETRATRIACONTANE	19.400	6687308	48.912 ug/ml
16) N-HEXATRIACONTANE	20.146	6689027	50.809 ug/ml
17) N-OCTATRIACONTANE	20.900	6169836	51.307 ug/ml
18) N-TETRACONTANE	21.830	6392299	51.910 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052419\
Data File : FF004628.D
Signal(s) : FID2B.ch
Acq On : 24 May 2019 19:04
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 24 22:50:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
Quant Title :
QLast Update : Thu May 23 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

