

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052819\
 Data File : FF004631.D
 Signal(s) : FID2B.ch
 Acq On : 28 May 2019 10:21
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: May 29 01:14:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
 Quant Title :
 QLast Update : Thu May 23 07:21:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.700	5686570	46.569 ug/ml
Target Compounds			
2) N-DECANE	4.419	7268438	49.637 ug/ml
3) N-DODECANE	6.442	7252722	49.011 ug/ml
4) N-TETRADECANE	8.238	7143130	48.629 ug/ml
5) N-HEXADECANE	9.840	7102876	48.108 ug/ml
6) N-OCTADECANE	11.284	6928021	47.719 ug/ml
7) N-EICOSANE	12.597	6848530	47.287 ug/ml
8) N-DOCOSANE	13.799	6609551	46.775 ug/ml
10) N-TETRACOSANE	14.905	6428028	46.383 ug/ml
11) N-HEXACOSANE	15.930	6320237	46.386 ug/ml
12) N-OCTACOSANE	16.882	6283445	46.215 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052819\
Data File : FF004631.D
Signal(s) : FID2B.ch
Acq On : 28 May 2019 10:21
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: May 29 01:14:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
Quant Title :
QLast Update : Thu May 23 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

