

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF053119\
 Data File : FF004670.D
 Signal(s) : FID2B.ch
 Acq On : 31 May 2019 16:21
 Operator : AJ\SJ
 Sample : K3105-09MSD
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: May 31 22:53:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
 Quant Title :
 QLast Update : Thu May 23 07:21:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.693	1690879	13.847 ug/ml
Target Compounds			
1) N-OCTANE	2.312	1699307	11.779 ug/ml
2) N-DECANE	4.415	2425028	16.561 ug/ml
3) N-DODECANE	6.436	2499278	16.889 ug/ml
4) N-TETRADECANE	8.231	2445620	16.649 ug/ml
5) N-HEXADECANE	9.833	2394868	16.220 ug/ml
6) N-OCTADECANE	11.278	2318897	15.972 ug/ml
7) N-EICOSANE	12.590	2291639	15.823 ug/ml
8) N-DOCOSANE	13.791	2216271	15.684 ug/ml
10) N-TETRACOSANE	14.897	2292722	16.544 ug/ml
11) N-HEXACOSANE	15.922	2635879	19.345 ug/ml
12) N-OCTACOSANE	16.875	3806498	27.997 ug/ml
13) N-TRIACONTANE	17.767	4387641	31.729 ug/ml
14) N-DOTRIACONTANE	18.603	3239177	23.416 ug/mlm
15) N-TETRATRIACONTANE	19.390	2157706	15.782 ug/ml
16) N-HEXATRIACONTANE	20.136	1957558	14.869 ug/mlm
17) N-OCTATRIACONTANE	20.886	1770760	14.725 ug/ml
18) N-TETRACONTANE	21.811	1842101	14.959 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF053119\
Data File : FF004670.D
Signal(s) : FID2B.ch
Acq On : 31 May 2019 16:21
Operator : AJ\SJ
Sample : K3105-09MSD
Misc :
ALS Vial : 59 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: May 31 22:53:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052219.M
Quant Title :
QLast Update : Thu May 23 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

