

Data Path : P:\HPCHEM1\FID_F\Data\FF060118\
 Data File : FF000718.D
 Signal(s) : FID2B.ch
 Acq On : 01 Jun 2018 23:40
 Operator : UA\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jun 02 05:29:26 2018
 Quant Method : P:\HPCHEM1\FID_F\Method\FF060118.M
 Quant Title :
 QLast Update : Fri Jun 01 16:01:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.491	5252971	45.550 ug/ml
Target Compounds			
2) N-DECANE	4.252	6547676	46.730 ug/ml
3) N-DODECANE	6.266	6487805	46.291 ug/ml
4) N-TETRADECANE	8.055	6424306	46.142 ug/ml
5) N-HEXADECANE	9.651	6344900	46.018 ug/ml
6) N-OCTADECANE	11.090	6254805	45.805 ug/ml
7) N-EICOSANE	12.397	6114831	45.641 ug/ml
8) N-DOCOSANE	13.592	6005585	45.419 ug/ml
10) N-TETRACOSANE	14.693	5892141	45.426 ug/ml
11) N-HEXACOSANE	15.713	5833785	45.419 ug/ml
12) N-OCTACOSANE	16.660	5780867	45.462 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : P:\HPCHEM1\FID_F\Data\FF060118\
Data File : FF000718.D
Signal(s) : FID2B.ch
Acq On : 01 Jun 2018 23:40
Operator : UA\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jun 02 05:29:26 2018
Quant Method : P:\HPCHEM1\FID_F\Method\FF060118.M
Quant Title :
QLast Update : Fri Jun 01 16:01:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

