

Method Path : P:\HPCHEM1\FID_F\Method\
 Method File : FF060118.M
 Title :
 Last Update : Fri Jun 01 16:01:51 2018
 Response Via : Initial Calibration

Calibration Files

50 =FF000699.D 100 =FF000698.D 20 =FF000700.D
 10 =FF000701.D 5 =FF000702.D

	Compound	50	100	20	10	5	Avg		%RSD
1)	N-OCTANE	1.428	1.391	1.364	1.395	1.468	1.409	E5	2.84
2)	N-DECANE	1.426	1.389	1.357	1.381	1.451	1.401	E5	2.68
3)	N-DODECANE	1.426	1.386	1.358	1.385	1.453	1.402	E5	2.69
4)	N-TETRADECANE	1.414	1.370	1.348	1.379	1.451	1.392	E5	2.90
5)	N-HEXADECANE	1.396	1.356	1.337	1.370	1.436	1.379	E5	2.77
6)	N-OCTADECANE	1.380	1.344	1.323	1.360	1.421	1.366	E5	2.74
7)	N-EICOSANE	1.351	1.320	1.300	1.331	1.397	1.340	E5	2.76
8)	N-DOCOSANE	1.330	1.299	1.284	1.313	1.386	1.322	E5	2.97
9) S	TETRACOSANE-d50 (S...	1.160	1.138	1.123	1.148	1.197	1.153	E5	2.44
10)	N-TETRACOSANE	1.306	1.278	1.261	1.290	1.351	1.297	E5	2.65
11)	N-HEXACOSANE	1.291	1.262	1.250	1.282	1.337	1.284	E5	2.61
12)	N-OCTACOSANE	1.273	1.248	1.237	1.271	1.328	1.272	E5	2.77
13)	N-TRIACONTANE	1.257	1.233	1.225	1.266	1.333	1.263	E5	3.37
14)	N-DOTRIACONTANE	1.215	1.193	1.189	1.235	1.302	1.227	E5	3.76
15)	N-TETRATRIACONTANE	1.193	1.175	1.166	1.202	1.265	1.200	E5	3.27
16)	N-HEXATRIACONTANE	1.136	1.123	1.109	1.144	1.197	1.142	E5	2.95
17)	N-OCTATRIACONTANE	1.024	1.017	0.996	1.026	1.082	1.029	E5	3.12
18)	N-TETRACONTANE	1.021	1.018	0.989	1.022	1.071	1.024	E5	2.86

(#) = Out of Range